






## Declaration of Conformity - CE

This Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER			
Name of Company	Address	SRN	Telephone/Email
Ferno-Washington, Inc. 	70 Weil Way, Wilmington, OH 45177-9371 U.S.A.	US-MF-000041603	+1 (937) 382-1451 fernoorders@ferno.com

EU AUTHORIZED REPRESENTATIVE			
Name of Company	Address	SRN	Telephone/Email
Ferno S.r.l. 	Via B. Zallone n.26 40066 Pieve di Cento (BO) Italy	IT-AR-000031265	+39 (051) 6860028 www.ferno.it

PRODUCT IDENTIFICATION	
Product Brand Name	Photo
POWER X2	
EMDN	
V08050102	
SELF-LOADING STRETCHERS	
Intended Purpose	
Power X2 Wheeled Stretchers are for professional use and are intended to be used as an emergency patient handling device and can be used to transport patients in ground ambulances.	

REF (Item / Catalog #)	Item Description	GTIN (UDI-DI)	GMN (Basic UDI-DI)
0015821	POWERX2 PREMIUM	00190790006600	0190790V08050102UL

Related Accessories (if applicable):

Item / Catalog #	Item Description	GTIN (UDI-DI)	GMN (Basic UDI-DI)
0567091	POWER F2 WITH POWERDRIVE (0567093 Carriage)	00190790006570	0190790V0880K5

RISK CLASS FOR MEDICAL DEVICES	
Device Classification	Common Specifications
Class I, Rule 1	Not applicable



According to:

HARMONIZED AND NON-HARMONIZED STANDARDS	
Item	Description
ISO 14971:2019	Medical devices – Application of risk management to medical devices
ISO 15223-1:2021	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements
ISO 13485:2016	Medical devices – Quality management systems – Requirements for regulatory purposes
EN 1865-2:2024	Patient Handling Equipment used in Road Ambulances – Part 2: Power Assisted Stretcher
EN 1865-3:2012+A1:2015	Patient Handling Equipment used in Road Ambulances – Part 3: Heavy Duty Stretcher
EN 1789:2020+A1:2023	Medical Vehicles and their Equipment – Road Ambulances

Complies with the essential requirements listed in Annex I of the Regulation (EU) 2017/745 concerning Medical Devices.

**COMPANY REPRESENTATIVE:**

DOROTHY RAMSEY

TITLE: PRRC, VP GLOBAL LEGAL & REGULATORY

SIGNATURE:

*Dorothy Ramsey*

PLACE: Wilmington, OH, USA

DATE: November 4, 2025

DoC Revision: 00

Record History:

Revision	Summary of Changes	Author	Effective Date
00	Initial issue	D. Greene	11/4/2025