



# AB GERMA®

## EU Declaration of conformity no. 210322-023

<b>Product Name:</b>	Positioning Pillow 55X30cm White
<b>Intended use:</b>	Germa Positioning Pillows are primarily intended to be used in prehospital, hospital, rehabilitation and dental treatment environment, by professional trained personnel, for a long and safe motion restriction and/ or positioning of body or extremities.
<b>SRN:</b> <b>Basic UDI-DI:</b> <b>UDI DI:</b> <b>Germa Article No:</b>	SE-MF-000003932 735001959P04POSPILLQC 07350019594407 22393300000
<b>Manufacturer:</b> <b>Visiting address:</b> <b>Phone:</b> <b>Email:</b> <b>Web:</b>	AB Germa Industrigatan 54-56, SE-29136 Kristianstad +46 (0)44 123030 <a href="mailto:info@germa.se">info@germa.se</a> <a href="http://www.germa.se">www.germa.se</a>
<b>Product class:</b>	Class I according to rule 1 in Annex VIII in MDR 2017/745
<b>Conformity procedure:</b>	Self-certification according to Annex IV in MDR 2017/745
<b>Identification:</b>	All products with serial numbers issued from; LOT number: 517470 Date: 2021-05-25

Declaration statement;

This EU declaration of conformity is issued under the sole responsibility of the manufacturer AB Germa. The devices covered by this declaration is in conformity with the requirements in the European Medical Device Regulation 2017/745.

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Managing Director  
Name: Björn Holmqvist  
Date: 2021-05-25

Sign: \_\_\_\_\_

Postaddress/Postal address	Adresal/Goods address	Telefon/Phone	Telefax/Fax	E-mail/Web page	VAT No.	Bankgiro Bank account
AB GERMA Box 2127 SE-291 02 KRISTIANSTAD SWEDEN	Industrigatan 54-56 SE-291 36 Kristianstad	+46 (0)44-123030	+46 (0)44-10 31 79	<a href="mailto:info@germa.se">info@germa.se</a> <a href="http://www.germa.se">www.germa.se</a>	SE556086941301	481-4126