



EC Declaration of conformity no.: 071004-002 (issue 2)

Product Name:	Set of Splints, type FERNO Washington UK
Ferno Article No:	AS 190
Germa Article No:	23005002210
Description:	<p>Vacuum splints are tools for fixation and immobilisation of patients with suspected injuries on arm, leg, shoulder or hip. They can also be used as vacuum mattresses for children.</p> <p>The splint consists of an airtight shell with fixation straps, vacuum valve and polystyrene granules.</p> <p>Set of Splints comprising of:</p> <ul style="list-style-type: none">1 x 23005002012, AS 100, Arm splint, short1 x 23005002013, AS 120, Arm splint1 x 23005002014, AS 140, Leg splint1 x 23005002026, AS 180, Bag for splints1 x 23005002017 Repair kit FW1 x 23272000711, Vacuum Hand Pump
Accessory:	

Identification: All products with serial numbers issued after LOT number: **030704**, date: **2003-07-04** (yyyy-mm-dd).

We the manufacturer hereby declare that this product is in conformity with the requirements in:

- Annex I and Annex VII and Annex IX in Council directive 93/42/EEC.
Equipment class: I, non-sterile, no measuring function.
- Swedish Medical Devices Act 1993:584

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Quality Manager
Name: Johan Attelid

Date: 19/12-2011

Sign: _____

Postadress	Adress	Telefon	Telefax	Organisationsnummer. F-skattebevis	Bankgiro	Postgiro
Box 2127 S-291 02 KRISTIANSTAD SWEDEN e-mail: info@germa.se	Industrigatan 54-56 KRISTIANSTAD	+46 (0)44-123030	+46 (0)44-103179	556086-9413	481-4125	21 68 22-7